

STUMPY Intra Uterine Insemination Cannula

Description:

STUMPY IUI Intra Uterine Insemination Cannula

Configuration:

See brochure

Indication for use:

Allwin Intra Uterine Insemination Catheters are intended for the introduction of washed spermatozoa into the uterine cavity.

Contraindications:

- Cervical infection
- Recent pelvic inflammation.
- Sexually transmitted disease
- Pregnancy
- Confirmed or suspected intrauterine device
- Uterine perforation.

Complications:

- Uterine wall perforation
- Bleeding

Caution:

- Federal (USA) law restricts this device to sale by or on the order of a physician.
- To reduce the risk of perforation, if any resistance is felt while inserting the catheter, do not force the catheter against the resistance.

Warning / Precaution:

- Do not use if package is opened or damaged.
- Not intended for IVF, GIFT or other Intra Fallopian tube procedure.
- Always use washed spermatozoa when performing intrauterine artificial insemination. The introduction of unwashed spermatozoa into the uterus may result in severe adverse reactions, including anaphylactic shock. Please refer to published medical literature for methods of preparing spermatozoa for intrauterine artificial insemination before performing the procedure.
- Dispose in accordance with recognized medical practice and under observance of applicable law and regulation.
- For single use only.
- Do not re-sterilize.
- Only experienced surgeon should use this device.
- Storage temperature between 5°C to 30°C.

Instruction for Use:

1. Before to the procedure, check the condition of the cervix prior to the insertion of the catheter, to ascertain the depth of the uterus and any uterine anteversion or retroflexion present in individual patients.
2. Place the patient in the dorsal lithotomy position with her feet in stirrups.
3. Position a vaginal speculum within the vagina so the external cervical os can be visualized during catheter insertion.

4. Cleanse the cervix with an appropriate cleansing solution.
5. Affix the adaptor of the catheter to a syringe preloaded with washed sperm solution.
6. Guide the catheter through the cervix and into the uterine cavity.
7. Pass the catheter into the uterine cavity until the bulb presses against the cervix.
8. Slowly press the plunger of the syringe to introduce the sperm into the uterine cavity.
9. Following injection, remove the catheter, discard the catheter, syringe, and other materials used.

NOTE-1 Human Sperm Survival Assay (HSSA) > 80% of control motility at 24 hours

NOTE-2 Endotoxin Testing: ≤ 20 EU/device

Do not reuse, reprocess & resterilize:

Reuse, Reprocessing & reesterilization may compromise the structural integrity, can also create risk of contamination & may not give desired result or create complications, infections which may result in injury, illness or death.

Limited Express Warranty:

The limited express warranty as set forth herein is exclusive and in lieu of all warranties of merchantability and fitness for use, remedies, obligations and liabilities for consequential damages. The products are being sold only for the purpose described herein and such limited express warranty runs only to the original user. In no event shall ALLWIN be liable for any breach of warranty in any amount exceeding the purchase price of the product. ALLWIN reserves the right to make design changes to products without liability to incorporate said changes in ALLWIN products previously designed or sold.

REF	Catalogue Number	LOT	Batch Code	Date of Manufacture	Use By	Do not re-use	Do not Re-Sterilize	Rx only	Caution: Federal Law restricts this device to sale by or on the order of a physician or a practitioner trained in its use.
Consult Instruction for use	Do not use if Packing is damaged	Caution	Keep out of sunlight	Keep Dry	STERILE	EO	Sterilized using ethylene oxide		

EC REP
CMC Medical Devices & Drugs S. L.
C/Horacio Lengua N° 18, CP 29006, Malaga, Spain
Tel.: +34 951 214 054
E-mail : info@cmcmedicaldevices.com



allwin®
Medical Devices
Allwin Medical Devices
Plot # 223, 230 & 221, Surat Special Economic Zone, Sachin,
Surat - 394230 Gujarat, INDIA | Tel: +91 99870 84561
E-mail : info@allwinmedical.com | www.allwinmedical.com