

# CURVED Intra Uterine Insemination Catheter

**Description:**

CODE	Description
I-UI-C	CURVED Intra Uterine Insemination Catheter
I-UI-C-O	CURVED Intra Uterine Insemination Catheter (Open Tip)

**Configuration:**

See brochure

**Indications for use:**

Allwin Intra Uterine Insemination Catheters are intended for the introduction of washed spermatozoa into the uterine cavity.

**Contraindications:**

- Cervical infection
- Recent pelvic inflammation.
- Sexually transmitted disease
- Pregnancy
- Confirmed or suspected intrauterine device
- Uterine perforation.

**Complications:**

- Uterine wall perforation
- Bleeding

**Caution:**

- Federal (USA) law restricts this device to sale by or on the order of a physician.
- To reduce the risk of perforation, if any resistance is felt while inserting the catheter, do not force the catheter against the resistance.

**Warning / Precaution:**

- Do not use if package is opened or damaged.
- Not intended for IVF, GIFT or other Intra Fallopian tube procedure.
- Always use washed spermatozoa when performing intrauterine artificial insemination. The introduction of unwashed spermatozoa into the uterus may result in severe adverse reactions, including anaphylactic shock. Please refer to published medical literature for methods of preparing spermatozoa for intrauterine artificial insemination before performing the procedure.
- Dispose in accordance with recognized medical practice and under observance of applicable law and regulation.
- For single use only.
- Do not re-sterilize.
- Only experienced surgeon should use this device.
- Storage temperature between 5°C to 30°C.

**Instruction for Use:**

1. Before to the procedure, check the condition of the cervix prior to the insertion of the catheter, to ascertain the depth of the uterus and any uterine ante flexion or retro flexion present in individual patients.
2. Place the patient in the dorsal lithotomy position with her feet in stirrups.
3. Position a vaginal speculum within the vagina so the external cervical os can be visualized during catheter insertion.
4. Cleanse the cervix with an appropriate cleansing solution.
5. Affix the adaptor of the catheter to a syringe preloaded with washed sperm solution.
6. Guide the catheter through the cervix and into the uterine cavity. **Note:** ink marks or depth positioners can be used to facilitate placement.
7. Slowly press the plunger of the syringe to introduce the sperm into the uterine cavity.
8. Following injection, remove the catheter, discard the catheter, syringe, and other materials used.

**NOTE-1** Human Sperm Survival Assay (HSSA) > 80%

**NOTE-2** Endotoxin Testing: ≤ 20 EU/device

**Do not reuse, reprocess & resterilize:**

Reuse, Reprocessing & reesterilization may compromise the structural integrity, can also create risk of contamination & may not give desired result or create complications, infections which may result in injury, illness or death.

**Limited Express Warranty:**

The limited express warranty as set forth herein is exclusive and in lieu of all warranties of merchantability and fitness for use, remedies, obligations and liabilities for consequential damages. The products are being sold only for the purpose described herein and such limited express warranty runs only to the original user. In no event shall ALLWIN be liable for any breach of warranty in any amount exceeding the purchase price of the product. ALLWIN reserves the right to make design changes to products without liability to incorporate said changes in ALLWIN products previously designed or sold.

Catalogue Number   
 Batch Code   
 Date of Manufacture   
 Use By   
 Do not re-use   
 Do not Re-Sterilize   
 **Rx only**   
 Caution: Federal Law restricts this device to sale by or on the order of a physician or a practitioner trained in its use.

Consult Instruction for use   
 Do not use if Packing is damaged   
 Caution   
 Keep out of sunlight   
 Keep Dry   
 MD Medical Device   
 STERILE EO Sterilized using ethylene oxide   
 Not made with natural rubber latex   
 Temperature Limit

CMC Medical Devices & Drugs S. L.  
 C/Horacio Lengo N° 18, CP 29006, Malaga, Spain  
 Tel.: +34 951 214 054  
 E-mail : info@cmcmedicaldevices.com

Single Sterile Barrier system with protective packaging inside

CE 0123

**allwin** Medical Devices  
**Allwin Medical Devices, Inc.**  
 3305 E. Miraloma Ave., Suite 176 Anaheim, CA 92806 (USA)  
 Tel. : +1 714-572-1709  
 Email : info@allwinmedical.com | www.allwinmedical.com